

**4<sup>TH</sup>**  
**2<sup>ND</sup>** **WFHSS**  
**CONGRESS**  
**BRUSSELS**  
18-21 OCTOBER  
2023

# A STUDY OF ESTABLISHMENT AND EVALUATION OF A RISK PREDICTION MODEL FOR STEAM STERILIZATION

**Name: Xin Zhao**

**Affiliation: Xuanwu Hospital Of Capital  
Medical University - Beijing. China**



首都医科大学宣武医院  
Xuanwu Hospital Capital Medical University

- **Established in 1958**
- **National Center for  
Neurological Disorders**
- **National Clinical Research  
Center for Geriatric Diseases**
- **Amount of Beds: 1643**
- **Amount of ORs: 39**




# background

- **Steam sterilization is still the most preferred method in hospital**
- **Quality control of sterilization process:**
  - Professional competence of staff
  - Performance of sterilizer

**“ A High level sterility assurance shall be achieved by  
effective combination of human and equipment! ”**

# background

## Sterilizer Unexpected Alarm



**Forced shutdown**



**Re- Sterilization**



**Delivery delay**



**Work overtime**



**Waste resources**

# background

- **24 alarms of sterilizers in total were triggered in 2021.**
- **320 packs had to be re-packed and re-sterilized due to those alarms.**


| Alarm                                   | Number of Alarm<br>(% of Total) | Resolution                                         |
|-----------------------------------------|---------------------------------|----------------------------------------------------|
| Alarm during <u>Pre-Vac</u> phase       | 4 (16%)                         | ① Pressure sensor calibration<br>② Tighten pipings |
| Alarm during <u>Sterilization</u> phase | 10 (42%)                        | Temperature sensor calibration                     |
| Alarm during <u>other</u> phases        | 10 (42%)                        | Replace PLC battery                                |

# Objective

- This experiment is an attempt to establish a sterilization risk prediction model, by applying criteria stricter than the control system of sterilizer, to proactively intervene in the sterilization process at an early stage thus provide "Early warning" of the sterilization quality.
- It eliminates the risk of failure much earlier and allows CSSD to manage the sterilizer in a proactive and predictable manner.



# Materials and Methods - Experiment Group



# Materials and Methods - A

## ➤ Object of evaluation

| Selected Sterilization Program | Moment of evaluation            | Parameter to be evaluated                                                                        |
|--------------------------------|---------------------------------|--------------------------------------------------------------------------------------------------|
| P1<br>(134°C, 5min)            | At completion of<br>every batch | T1: Control temperature<br>T2: Record temperature<br>P1: Control pressure<br>P2: Record pressure |



# Materials and Methods - A

Step 1: Read batch report carefully

```
Signature:
Cycle approved: V/N Date:
CYCLE PASSED

FB value 136.8min
Time Drain above ster. temp. 5:00m:s
Max. sterilize temperature 135.1°C
Min. sterilize temperature 134.4°C

53:38 Complete 953 65.6
52:19 Air Break 58 73.4
49:19 Dry 91 59.7
47:15 Exhaust 781 64.5
45:44 Hold Air 796 70.8
44:49 Air Break 68 76.7
41:49 Dry 98 88.9
39:43 Exhaust 786 68.2
38:12 Hold Air 798 71.3
37:16 Air Break 58 74.6
26:16 Dry 91 59.3
22:08 Exhaust 3109 134.9
22:07 3111 134.9
21:07 3094 134.8
20:07 3096 134.8
19:07 3101 134.9
18:07 3124 134.8
17:07 Sterilize 3102 134.3
13:18 Heat-up 73 71.7
10:19 3. Vacuum pulse 1822 116.8
8:38 2. Steam pulse 78 52.9
5:38 2. Vacuum pulse 1888 114.7
3:41 1. Steam pulse 62 62.1
0:01 1. Vacuum pulse 998 45.9

#s
Time Phase : 22.08.2023 / 13:31 mbara T2 °C
Cycle start : 17.0min
Dry time : 134.2°C
Sterilize temp. : 5.8min
Sterilize time : 3
No. of pulses :

Set points : PR: 05.06.2008 SW: U.5.1
: 1: Instruments 134°C
Operator : User
Cycle counter : 000063
Machine type : 9-6-15 HS2 No:2005375
Department : CSSD
Hospital : XAVY-4
BELIED CYCLE DOCUMENTATION
```

Step 2: Understand alarm criteria

| Phase         | Machine Alarm Criteria                                                                                                      |
|---------------|-----------------------------------------------------------------------------------------------------------------------------|
| Pre-Vac       | Vac Time > <b>15</b> min                                                                                                    |
| Sterilization | 1 Control Temp T1 < <b>134</b> °C<br>2 Deviation between T1&T2 > <b>1</b> °C<br>3 Deviation between P1&P2 > <b>100</b> mbar |
| Other         | PLC battery running $\leq$ <b>24</b> months                                                                                 |

Step 3: Determine preventive intervention criteria

| Phase         | Preventive Intervention Criteria                                                                                        |
|---------------|-------------------------------------------------------------------------------------------------------------------------|
| Pre-Vac       | Vac Time $\geq$ <b>8</b> min                                                                                            |
| Sterilization | 1 T1 or T2 < <b>134.2</b> °C<br>2 Deviation between T1&T2 > <b>0.6</b> °C<br>3 Deviation between P1&P2 > <b>60</b> mbar |
| Other         | PLC battery running $\leq$ <b>12</b> months                                                                             |

# Materials and Methods - A

|                               | Control Group                                                  | Experiment Group                                                                                                                                                                                                                   |
|-------------------------------|----------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Period</b>                 | 2021.8---2022.4 (9 Months)                                     | 2022.5---2023.1 (9 Months)                                                                                                                                                                                                         |
| <b>Batches</b>                | 4115                                                           | 4142                                                                                                                                                                                                                               |
| <b>Evaluation and Actions</b> | Evaluate batch report against EN285                            | <ul style="list-style-type: none"> <li>Evaluate batch report against EN285</li> <li>Implement the new preventive intervention criteria</li> </ul>                                                                                  |
|                               | Only contact service whenever an unexpected alarm is triggered | <ul style="list-style-type: none"> <li>Contact service whenever one of the preventive intervention criteria was reached</li> <li>Sterilizer still runs normally while waiting for preventive service action to be taken</li> </ul> |

# Intermediate Results of Study

| #        | Pre-Vac Phase | Sterilization Phase | Batch No. ( Sterilizer No.) | Criteria triggered Preventive Intervention | Preventive Service Action | Time for Preventive Service |
|----------|---------------|---------------------|-----------------------------|--------------------------------------------|---------------------------|-----------------------------|
| 1        | √             |                     | 752359 (#2)                 | 1st Vac Pulse >8min                        | Replace hose connection   | 15min                       |
| 2        | √             |                     | 3011626 (#4)                | 1st Vac Pulse >8min                        | Calibrate P sensor        | 35min                       |
| 3        | √             |                     | 3011626 (#4)                | 1st Vac Pulse >8min                        | Tighten hose connection   | 5min                        |
| <b>4</b> |               | √                   | <b>751357 (#3)</b>          | <b>T1 &lt;134.2°C (T1= 134.1°C)</b>        | <b>Calibrate T sensor</b> | <b>40min</b>                |
| 5        |               | √                   | 751357 (#3)                 | Deviation T1&T2 >0.6°C                     | Calibrate T sensor        | 45min                       |
| 6        |               | √                   | 752329 (#1)                 | Deviation T1&T2 >0.6°C                     | Calibrate T sensor        | 40min                       |
| 7        |               | √                   | 752329 ( #1)                | T2 <134.2°C (T2= 134.1°C)                  | Calibrate T sensor        | 45min                       |

**Totally 3h25min**

# Intermediate Results of Study - A

3-2

Signature: \_\_\_\_\_

Cycle approved: With Dates: \_\_\_\_\_

CYCLE PASSED *AWJ*

|                              |          |
|------------------------------|----------|
| F0 value                     | 129.1min |
| Time Drain above ster. temp. | 5:00m:5  |
| Max. sterilize temperature   | 134.9°C  |
| Min. sterilize temperature   | 134.1°C  |

|                       |      |       |
|-----------------------|------|-------|
| 66:18 Complete        | 949  | 55.4  |
| 64:59 Air Break       | 59   | 58.5  |
| 61:58 Dry             | 94   | 65.0  |
| 59:58 Exhaust         | 802  | 56.3  |
| 58:19 Hold Air        | 797  | 58.7  |
| 57:24 Air Break       | 59   | 61.3  |
| 54:23 Dry             | 95   | 67.5  |
| 52:11 Exhaust         | 806  | 60.9  |
| 50:40 Hold Air        | 797  | 63.0  |
| 49:44 Air Break       | 56   | 65.5  |
| 39:47 Dry             | 95   | 87.7  |
| 31:50 Exhaust         | 3879 | 134.9 |
| 30:50                 | 3872 | 134.9 |
| 29:50                 | 3871 | 134.8 |
| 28:50                 | 3879 | 134.9 |
| 27:50                 | 3874 | 134.9 |
| 26:49 Sterilize       | 3868 | 134.8 |
| 22:42 Heat-up         | 76   | 71.0  |
| 14:43 3. Vacuum pulse | 1881 | 116.2 |




Case #4  
T1 < 134.2°C (T1 = 134.1°C)

|                              |          |
|------------------------------|----------|
| F0 value                     | 129.1min |
| Time Drain above ster. temp. | 5:00m:5  |
| Max. sterilize temperature   | 134.9°C  |
| Min. sterilize temperature   | 134.1°C  |





# Methods Optimization 2 - C



---

• **SVSM% = ABS (Target Value – Measured value) / Target value**

---

① Sterilization Temperature Band SVSM% =  $(3^{\circ}\text{C} - \text{Measured Sterilization Temperature Band}) / 3^{\circ}\text{C} \times 100\%$

---

② Holding Time SVSM% =  $(\text{Measured Holding Time} - 180\text{s}) / 180\text{s} \times 100\%$

---

③ Sterilization Temperature Deviation SVSM % =  $(2^{\circ}\text{C} - \text{Sterilization Temperature Deviation}) / 2^{\circ}\text{C} \times 100\%$

---

• **SPSM% = Min ( all above 3 SVSM%)**

---


# Results & Discussions

➤ In total 8257 batches have been evaluated

| Basic Fact of Study                                                      | Control Group | Experiment Group |
|--------------------------------------------------------------------------|---------------|------------------|
| Number of batches (134°C,5min)                                           | 4115          | 4142             |
| Number of preventive interventions under<br><b>Risk Prediction Model</b> | N/A           | 7                |

# Results & Discussions

## ➤ Qualification of Sterilization Pack

| Indicator of Result                       | Control Group | Experiment Group | Improvement                                                                                 |
|-------------------------------------------|---------------|------------------|---------------------------------------------------------------------------------------------|
| Number of sterilization packs processed   | 155664        | 146183           |                                                                                             |
| Number of unqualified sterilization packs | 315           | 0                |                                                                                             |
| Qualification rate of sterilization packs | 99.78%        | 100%             | 0.22%  |



# Results & Discussions


## ➤ Sterilizer Operation Efficiency

| Indicator of Result                                                     | Control Group | Experiment Group | Improvement |
|-------------------------------------------------------------------------|---------------|------------------|-------------|
| Operation Time (h) (A)                                                  | 3925.82       | 3489.70          |             |
| <b>Proactive</b> shutdown due to preventive intervention service(h) (B) | 0             | 3.41             |             |
| <b>Passive</b> shutdown due to unexpected alarm and service(h) (C)      | 339.93        | 0                |             |
| Rate of sterilizer proper operation%* (D)                               | 91.34%        | 99.90%           | 8.56% ↑     |

$$D = (A - B - C) / A \times 100\%$$

# Results & Discussions

## ➤ CSSD Work Efficiency



| Indicator of Result                                              | Control Group | Experiment Group | Improvement |
|------------------------------------------------------------------|---------------|------------------|-------------|
| Percentage of delivery delay due to sterilizer unexpected alarm* | 59%           | 0                | 59% ↓       |
| OT due to sterilizer unexpected alarm (h)                        | 279.50        | 0                |             |
| Percentage of OT due to sterilizer unexpected alarm %            | 37%           | 0                | 37% ↓       |

Breakdown of CSSD delivery delay (Control Group)

\*Percentage of delivery delay due to sterilizer alarm% =

$$\frac{\text{Delivery delay cases due to sterilizer alarm}}{\text{Total delivery delay cases}} \times 100\%$$

# Results & Discussions

- We invited Dr. Zhang Jinxin and his team from Sun Yat-sen University to analyse two groups' data generated by 4 sterilizers by statistical method



## Possible Explanation

- Tolerance of parameter already rather small
- Machine performance quite stable

| Indicators Analysed                                            | EN285  | P Value |
|----------------------------------------------------------------|--------|---------|
| Temperature deviation during sterilization phase (T1,T2,T3,T4) | < 2°C  | > 0.05  |
| Temperature fluctuation during sterilization phase             | < 3°C  | > 0.05  |
| Holding time                                                   | > 180s | > 0.05  |

# Results & Discussions

➤ Improvement of sterilization performance of one sterilizer is observed


| #1 Sterilizer                               | min<br>T1/T2/T3/T4<br>°C | max<br>T1/T2/T3/T4<br>°C | T Deviation<br>°C        | Rate of T Deviation<br>(Deviation/2°C) | T Fluctuation<br>°C   | Rate of T Fluctuation<br>(Fluctuation/ 3°C) |
|---------------------------------------------|--------------------------|--------------------------|--------------------------|----------------------------------------|-----------------------|---------------------------------------------|
|                                             | Temperature Precision    |                          | Temperature Distribution |                                        | Temperature Stability |                                             |
| Without new model<br>(740 batches in total) | 134.2                    | 135.2                    | 0.6                      | 29%                                    | 1.0                   | 33%                                         |
| Under new model<br>(685 batches in total)   | <b>134.4</b>             | <b>135.0</b>             | <b>0.5</b>               | <b>26%</b>                             | <b>0.6</b>            | <b>21%</b>                                  |
| <b>Improvement</b>                          | <b>0.2°C</b> ↑           | <b>0.2°C</b> ↓           | <b>0.1°C</b> ↓           | <b>3%</b> ↓                            | <b>0.4°C</b> ↓        | <b>12%</b> ↓                                |

# Results & Discussions



Case #4

T1 < 134.2°C (T1= 134.1°C)




Was the root cause correctly identified and removed by this service intervention?



# Results & Discussions



**Case #7**  
T2 < 134.2°C (T2= 134.1°C)



Introducing more process parameters from the batch documentation system could give more precision to the Risk Preventive Model, and at the same time help to create a new quantitative tool for assessing the effect of every service intervention



# Conclusion

**1** New Management Approach



**2** Operate more independently

**3** Feasibility and Promotion

**4** Optimize Continuously - **D**



# Acknowledgement



- Ms Liu Ting, Head nurse of OR and CSSD.  
(Xuanwu Hospital Of Capital Medical University)
- Dr. Zhang Jinxin and his team  
(Sun Yat-Sen University )
- CSSD branch of Chinese Nursing Association



**24**  
WFHSS  
CONGRESS  
**BRUSSELS**  
18-21 OCTOBER  
2023



**Thanks for your attention!**